

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

Key to Log:

File Name: Name of excel files containing the data for that month. File names have two or three parts: 1) the two-letter site abbreviation (GJ, HB, BA, etc.), 2) the year and month (ie. -0201), and 3) the nature of the file (“-raw” contains the raw, unaltered data; “-QAQC” contains the quality controlled data set as well as all corrections; the file name that ends with the year and month is the quality controlled file containing only the corrected and finalized data—this is the file sent to the archive)

Deployments: Number of different sondes that recorded data during the month and the periods of dates/times of each deployment.

Condition of Sonde: The post-deployment condition of each sonde deployed during the month. This includes information on fouling, equipment failures and whether post-deployment checks were performed.

Removed Data: Tabulation of all data points removed from a given month. “Trimming on ends of data sets” is a record of all data points removed from either the beginning or the end of the different files in order to create a seamless monthly record (most points removed here were data not recorded in the water, but rather, were point recorded prior to deployment or following retrieval); “Removal of bad data” is a record of data deemed to be of low quality (for example, data out of range of instrument, instrument or probe failures, etc... See Word file “QAQCGuidelines.doc” for criteria used). Table columns give the parameter values deleted, the reason for the deletion (see abbreviations) and the dates and times of points deleted.

Corrected data: This is a record of all data points that were corrected. This includes corrections due to instrument drift, fouling, incorrect instrument calibration, etc. Included are probe readings in the standard pre- and post-deployment and excel formulae used to calculate corrected values. Inability to correct data due to lack of proper post-deployment check procedures or substandard sonde condition (eg. heavily fouled) may also be noted here.

Missing data: This is a record of all missing data points not due to the QA/QC process (ie. not accounted for in “Removal of bad data”). A common cause for this missing data is a lag time between the retrieval of one sonde and the deployment of the second sonde or failure of the instrument to log data at a given time.

Problems and Anomalies: This is a record of troublesome trends or data points not removed from data set, but that could prove a problem in interpretation. Examples include sudden jumps in the data when sondes are changed out (reflecting drift in retrieved sonde or a lack of standardization between the two sondes). Notes regarding reliability of data (whether or not it is or may be faulty) may also be found here. **ALWAYS read this section before interpreting data.**

Abbreviations:

IF	=	Instrument Failure: Data logger returned values of -6999
PF	=	Probe Failure: Probe measuring individual parameter apparently malfunctioned.
ADL	=	Above Detection Limit: data logger returned a data point that is above the detection limit of the probe
BDL	=	Below Detection Limit: data logger returned a data point that is above the detection limit of the probe
SND	=	Sonde Not Deployed: evidence indicates that sonde was not in the water on-site when data was recorded
FOUL	=	Fouled: evidence indicates sonde was not functioning properly due to severe fouling
EXP	=	Exposed: Sonde was exposed to air due to low water level or some disturbance.
NMD	=	Next month's data: trimmed data belonged to next month
PMD	=	Previous month's data: trimmed data belonged to previous month

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

General Notes on Reliability of Data:

- 1) In general, measurements of temperature and depth are very reliable unless otherwise noted in "Problems and Anomalies".
- 2) Salinity is typically reliable, but this data can be compromised by bad calibrations and fouling. These effects are most obvious as sudden discontinuities in the trend when sondes are changed. If the discontinuity that occurs with a sonde change is more than +/- 2 ppt in magnitude, the discontinuity is noted as a faulty trend.
- 3) Measurements of dissolved oxygen are often not reliable. Typically, oxygen measurements taken soon after a sonde is deployed are reliable, but reliability decreases during the deployment period due to instrument drift and fouling. The most unreliable oxygen data is that collected near the time the sonde is retrieved. ALWAYS read "Problems and Anomalies" before interpreting dissolved oxygen! Dissolved oxygen discontinuities of +/- 25% or more coincident with sonde changes are noted as faulty if they do not fall within the actual rate of change occurring before and after the sonde change.
- 4) The reliability of turbidity measurements is much like that of oxygen. Turbidity measurements are best early and worst late in the deployment period.
- 5) The reliability of chlorophyll measurements is unknown. We do not currently know what the measurements mean in a biological context. Confirmation studies are underway.
- 6) If the word "**faulty**" appears regarding a trend or data period, the data should be considered highly unreliable. Do not use this data (if it wasn't deleted altogether) for anything but a general guideline to potential conditions. This designation is only used regarding data known to be of very poor quality.
- 7) If the phrase "**may be faulty**" appears regarding a trend or data period, the data may not be reliable. Typically, the data appears to be of reasonably good quality and probably does reflect the real trends in environmental condition, but very strict interpretation is not recommended.
If a proper post-deployment check was not performed, reliability of all data for that deployment period must be considered suspect.

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

JULY-2003

Files: Data: DI-0307-raw, DI-0307-QAQC, DI-0307

Deployments: (5); 6/26-7/3, 7/3-7/10, 7/10-7/18, 7/18-7/28, 7/28-8/6

Condition of Sondes: 6/26-7/3 (no depth probe; moderate fouling), 7/3-7/10 (heavy fouling), 7/10-7/18 (moderate fouling), 7/18-7/28 (no depth probe; very heavy fouling), 7/28-8/6 (very heavy fouling)

Removed Data:

Parameter(s)	Problem	Data Points
All parameters	EXP	6/26 1830-2000, 6/27 1900-2030, 6/29 2030-2230, 7/2 2330
All parameters	EXP	7/4 000, 7/9 1700-1800
All parameters	EXP	7/10 1730-2000, 7/11 1900-2030, 7/12 1930-2200, 7/13 2030 2230, 7/14 2100-2330, 7/15 2230-2330, 7/16 2300-2330
Turbidity	FOUL	7/17 100-7/18
Chlorophyll	FOUL	7/18 100-7/18
All parameters	EXP	7/26 1900-2100, 7/27 1930-2200
Oxygen	FOUL	7/24 030-7/28
Turbidity	FOUL	7/24 2200-7/28
Chlorophyll	FOUL	7/26 830-7/28
All parameters	EXP	7/28 2000-2200, 7/29 2130-2230, 7/30 2130-2330, 7/31 2200-2330, 8/1 2330
Oxygen	FOUL	8/1 1800-8/6
Chlorophyll	FOUL	8/4 1300-8/6

Corrected Data:

6/26-7/3

Specific Conductivity:

standard 10, probe 9.14

Formula: $= -((9.14-10)/(\$B\$337-\$B\$2))* (B2-\$B\$337)+D2$

Salinity $= (0.68*R2)-1.89$

Oxygen:

pre-deployment O2: 97.5

post-deployment O2: 100.6

standard 100, probe 94.8

Formula: $= -((100.6-97.5)/(\$B\$337-\$B\$2))* (B2-\$B\$2)+F2 + (100-97.5)$

Conversion for O2 concentration: $= 3.7 - (0.104*C2) - (0.0344*S2) + (0.0688*P2)$

Turbidity:

standard 0, probe 15.7; standard 123, probe 109.2

Formula: $= (((((123/93.5)-1)*(J2)) - (15.7))* (B2-\$B\$2)/(\$B\$337-\$B\$2)) + J2$

Chlorophyll:

standard 0.0, probe 4.1

Formula: $= -((4.1-0)/(\$B\$337-\$B\$2))* (B2-\$B\$2)+K2$

7/3-7/10

Specific Conductivity:

standard 12.88, probe 13.36

Formula: $= -((85.4-101.3)/(\$B\$673-\$B\$338))* (B338-\$B\$338)+F338 + (100-101.3)$

Salinity $= 3.85 - (0.103*C338) - (0.0313*S338) + (0.0662*P338)$

Oxygen:

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

pre-deployment O2: 101.3
post-deployment O2: 85.4
standard 100, probe 50.5
Formula: $= -((13.36 - 12.88) / (\$B\$673 - \$B\$338)) * (B338 - \$B\$338) + D338$
Conversion for O2 concentration: $= (0.699 * R338) - 2.58$

Turbidity:

standard 0, probe 2.6; standard 123, probe 72.9
Formula: $= (((((123/70.3) - 1) * (J338)) - (2.6)) * ((B338 - \$B\$338) / (\$B\$673 - \$B\$338))) + J338$

Chlorophyll:

standard 0.0, probe 3.1
Formula: $= -((3.1 - 0) / (\$B\$673 - \$B\$338)) * (B338 - \$B\$338) + K338$

7/10-7/18

Specific Conductivity:

standard 12.88, probe 12.61
Formula: $= -((12.61 - 12.88) / (\$B\$1050 - \$B\$674)) * (B674 - \$B\$674) + D674$
Salinity $= (0.678 * R674) - 1.85$

Oxygen:

pre-deployment O2: 97.6
post-deployment O2: 99.8
standard 100, probe 99.8
Formula: $= -((99.8 - 97.6) / (\$B\$1050 - \$B\$674)) * (B674 - \$B\$674) + F674 + (100 - 97.6)$
Conversion for O2 concentration: $= 3.84 - (0.107 * C674) - (0.0315 * S674) + (0.0676 * P674)$

Turbidity:

standard 0, probe 66.9; standard 123, probe 159.2
**No correction possible due to fouling

Chlorophyll:

standard 0.0, probe 16.1
**No correction possible due to fouling

7/18-7/28

Specific Conductivity:

standard 12.88, probe 11.02
Formula: $= -((11.02 - 12.88) / (\$B\$1540 - \$B\$1051)) * (B1051 - \$B\$1051) + D1051$
Salinity $= (0.694 * R1051) - 2.37$

Oxygen:

pre-deployment O2: 100
post-deployment O2: 12.9
standard 100, probe 9.1
**No correction possible due to fouling

Turbidity:

standard 0, probe 96; standard 123, probe 150.9
**No correction possible due to fouling

Chlorophyll:

standard 0.0, probe 18.4
**No correction possible due to fouling

7/28-8/6

Specific Conductivity:

standard 12.88, probe 11.6
Formula: $= -((11.6 - 12.88) / (\$B\$430 - \$B\$2)) * (B2 - \$B\$2) + D2$
Salinity $= (0.6963 * R2) - 2.4414$

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

Oxygen:

pre-deployment O2: 111.3
post-deployment O2: 14.6
standard 100, probe 10.6
**No correction possible due to fouling

Turbidity:

standard 0, probe 9.6; standard 100, probe 121.8
Formula: $=((((123/121.8)-1)*(J2))-(9.6))*((B2-\$B\$2)/(\$B\$430-\$B\$2))+J2$

Chlorophyll:

standard 0.0, probe 3.9
Formula: $=(-(3.9-0)/(\$B\$430-\$B\$2))*(B2-\$B\$2))+K2$

Problems and Anomalies:

All parameters numerous data points: Exposed at low tide. Points reflecting exposure of the sonde to air during low tide were **faulty** and were deleted.

Depth 6/26-7/3: The depth probe was not working during this period and no data was recorded.

Turbidity 7/17 100-7/18: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted. The remaining uncorrected portion of the data record **may be faulty** due to drift and/or fouling.

Chlorophyll 7/18 100-7/18: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted. The remaining uncorrected portion of the data record **may be faulty** due to drift and/or fouling.

Depth 7/18-7/28: The depth probe was not working during this period and no data was recorded.

Oxygen 7/24 030-7/28: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted. The remaining uncorrected portion of the data record **may be faulty** due to drift and/or fouling.

Turbidity 7/24 2200-7/28: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted. The remaining uncorrected portion of the data record **may be faulty** due to drift and/or fouling.

Chlorophyll 7/26 830-7/28: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted. The remaining uncorrected portion of the data record **may be faulty** due to drift and/or fouling.

Oxygen 8/1 1800-8/6: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted. The remaining uncorrected portion of the data record **may be faulty** due to drift and/or fouling.

Chlorophyll 8/4 1300-8/6: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

AUGUST--2003

Files: Data: DI-0308-raw, DI-0308-QAQC, DI-0308

Deployments: (4); 7/28-8/6, 8/6-8/15, 8/15-8/27, 8/27-9/11

Condition of Sondes: 7/28-8/6 (very heavy fouling), 8/6-8/15 (no data), 8/15-8/27 (very heavy fouling), 8/27-9/11 (very heavy fouling)

Removed Data:

Parameter(s)	Problem	Data Points
All parameters	EXP	7/28 2000-2200, 7/29 2130-2230, 7/30 2130-2330, 7/31 2200-2330, 8/1 2330
Oxygen	FOUL	8/1 1800-8/6
Chlorophyll	FOUL	8/4 1300-8/6
All parameters	EXP	8/24 1830-2200, 8/25 2000-2100, 8/26 2000-2130, 8/27 2100-2130
Oxygen	FOUL	8/23 1930-8/27
Chlorophyll	FOUL	8/23 1930-8/27
All parameters	EXP	8/28 2200, 9/8 2000-2030, 9/9 2030-2100, 9/10 2100-2130
Oxygen	FOUL	9/3 2200-9/11
Chlorophyll	FOUL	9/10 1930-9/11
Turbidity	FOUL	9/10 2200-9/11

Corrected Data:

7/28-8/6

Specific Conductivity:

standard 12.88, probe 11.6

Formula: $= -((11.6 - 12.88) / (\$B\$430 - \$B\$2)) * (B2 - \$B\$2) + D2$

Salinity $= (0.6963 * R2) - 2.4414$

Oxygen:

pre-deployment O2: 111.3

post-deployment O2: 14.6

standard 100, probe 10.6

**No correction possible due to fouling

Turbidity:

standard 0, probe 9.6; standard 100, probe 121.8

Formula: $= (((((123 / 121.8) - 1) * (J2)) - (9.6)) * ((B2 - \$B\$2) / (\$B\$430 - \$B\$2))) + J2$

Chlorophyll:

standard 0.0, probe 3.9

Formula: $= -((3.9 - 0) / (\$B\$430 - \$B\$2)) * (B2 - \$B\$2) + K2$

8/6-8/15

**No Data

8/15-8/27

Specific Conductivity:

standard 12.88, probe 12.1

Formula: $= -((12.1 - 12.88) / (\$B\$1428 - \$B\$865)) * (B865 - \$B\$865) + D865$

Salinity $= (0.6639 * R865) - 1.434$

Oxygen:

pre-deployment O2: 104.8

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

post-deployment O2: 32
standard 100, probe 35.2
**No correction possible due to fouling

Turbidity:

standard 0, probe -0.3; standard 100, probe 117.9
Formula: $=((((123/117.8)-1)*(J865))-0.4)*((B865-\$B\$865)/(\$B\$1428-\$B\$865))+J865$

Chlorophyll:

standard 0.0, probe 2.2
**No correction possible due to fouling

8/27-9/11

Specific Conductivity:

standard 10, probe 13.11
Formula: $=-((13.11-10)/(\$B\$729-\$B\$2))*(B2-\$B\$2)+E2$
Salinity $=(0.6808*R2)-1.903$

Oxygen:

pre-deployment O2: 99.225
post-deployment O2: 48.85
**No correction possible due to fouling

Turbidity:

standard 0, probe -0.3; standard 123, probe 117.9
Formula: $=((((123/118.2)-1)*(K2))-(-0.3))*((B2-\$B\$2)/(\$B\$729-\$B\$2))+K2$

Chlorophyll:

standard 0.0, probe 2.2
Formula: $=-(((2.2-0)/(\$B\$729-\$B\$2))*(B2-\$B\$2))+L2$

Problems and Anomalies:

All parameters numerous data points: Exposed at low tide. Points reflecting exposure of the sonde to air during low tide were **faulty** and were deleted.

Depth 8/15-8/27: The depth probe was not working during this period and no data was recorded.

Oxygen 8/1-8/6: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted. The remaining uncorrected portion of the data record **may be faulty** due to drift and/or fouling.

Chlorophyll 8/4-8/6: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Oxygen 8/23-8/27: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted. The remaining uncorrected portion of the data record **may be faulty** due to drift and/or fouling.

Chlorophyll 8/23-8/27: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted. The remaining uncorrected portion of the data record **may be faulty** due to drift and/or fouling.

Oxygen 9/3-9/11: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted. The remaining uncorrected portion of the data record **may be faulty** due to drift and/or fouling.

Chlorophyll 9/11: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Turbidity 9/11: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

SEPTEMBER--2003

Files: Data: DI-0309-raw, DI-0309-QAQC, DI-0309

Deployments: (4); 8/27-9/11, 9/11-9/18, 9/18-9/26, 9/26-10/10

Condition of Sondes: 8/27-9/11 (very heavy fouling), 9/11-9/18 (very heavy fouling), 9/28-9/26 (very heavy fouling), 9/26-10/10 (very heavy fouling caused probe malfunctions)

Removed Data:

Parameter(s)	Problem	Data Points
All parameters	EXP	Numerous
Oxygen	FOUL	9/3 2200-9/11 1430
Oxygen	FOUL	9/16 1730 – 9/18 1330
Oxygen	FOUL	9/23 300-9/26 1500
Oxygen	FOUL	10/4 2030-10/10 1300
Turbidity	FOUL	9/11 330-9/11 1430
Turbidity	FOUL	9/25 1830-9/26 1500
Turbidity	FOUL	10/8 1330-10/10 1300
Chlorophyll	FOUL	10/8 1330-10/10 1300

Corrected Data:

8/27-9/11

Specific Conductivity:

standard 10, probe 13.11

Formula: $= -(((13.11-10)/(\$B\$729-\$B\$2)) * (B2-\$B\$2)) + E2$

Salinity $= (0.6808 * R2) - 1.903$

Oxygen:

pre-deployment O2: 99.225

post-deployment O2: 48.85

**No correction possible due to fouling

Turbidity:

standard 0, probe -.3; standard 123, probe 117.9

Formula: $= ((((((123/118.2) - 1) * (K2)) - (-0.3)) * ((B2 - \$B\$2) / (\$B\$729 - \$B\$2)))) + K2$

Chlorophyll:

standard 0.0, probe 2.2

Formula: $= -(((2.2-0)/(\$B\$729-\$B\$2)) * (B2-\$B\$2)) + L2$

9/11-9/18

Specific Conductivity:

standard 10, probe na

**No correction possible

Oxygen:

pre-deployment O2: 98.18

post-deployment O2: 53.1

**No correction possible due to fouling

Turbidity:

standard 0, probe -.6; standard 123, probe 71.8

**probe malfunctioning

Chlorophyll:

standard 0.0, probe 0.6

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

Formula: $=(-((0.6-0)/(\$B\$1063-\$B\$730))*(B730-\$B\$730))+L730$

9/18-9/26

Specific Conductivity:

standard 12.88, probe 12.55

Formula: $=(-((12.55-12.88)/(\$B\$1446-\$B\$1064))*(B1064-\$B\$1064))+E1064$

Salinity $=(0.7058*R1064)-2.7685$

Oxygen:

pre-deployment O2: 102.4

post-deployment O2: 45.3

**No correction possible due to fouling

Turbidity:

standard 0, probe 35.8; standard 123, probe 115.6

**No correction possible due to fouling

Chlorophyll:

standard 0.0, probe 1.7

Formula: $=(-((1.7-0)/(\$B\$1446-\$B\$1064))*(B1064-\$B\$1064))+L1064$

9/26-10/10

Specific Conductivity:

standard 12.88, probe 11.8

Formula: $=(-((11.8-12.88)/(\$B\$2113-\$B\$1447))*(B1447-\$B\$1447))+E1447$

Salinity $=(0.7098*R1447)-2.8635$

Oxygen:

pre-deployment O2: 100.5

post-deployment O2: 6.67

**No correction possible due to fouling

Turbidity:

standard 0, probe 241.2; standard 100, probe 106.3

**No correction possible due to fouling

Chlorophyll:

standard 0.0, probe 44.5

**No correction possible due to fouling

Problems and Anomalies:

Oxygen 9/3-9/11: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Turbidity 9/11: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Salinity 9/11: Salinity increased from 20ppt to 26.5ppt when the sondes were changed. This discontinuity is **faulty**. Notice that during all the deployment periods, salinity slowly decreases then suddenly increases when a new sonde is deployed. This likely resulted from fouling slowly depressing the salinity measurements by the probe over the course of the deployment period.

Depth 9/18-9/26: There was not a depth probe on the sonde during this period.

Oxygen 9/16-9/18: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Salinity 9/18: Salinity increased from 21.6ppt to 24.2ppt when the sondes were changed. This discontinuity is **faulty**. Notice that during all the deployment periods, salinity slowly decreases then suddenly increases when a new sonde is deployed. This likely resulted from fouling slowly depressing the salinity measurements by the probe over the course of the deployment period.

Oxygen 9/23-9/26: Probe was heavily fouled resulting in readings that were **faulty** and were not correctable. The affected portion of the data was deleted.

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

Turbidity 9/25-9/26: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Salinity 9/26: Salinity increased from 23.6ppt to 26ppt when the sondes were changed. This discontinuity is **faulty**. Notice that during all the deployment periods, salinity slowly decreases then suddenly increases when a new sonde is deployed. This likely resulted from fouling slowly depressing the salinity measurements by the probe over the course of the deployment period.

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

OCTOBER--2003

Files: Data: DI-0310-raw, DI-0310-QAQC, DI-0310

Deployments: (3); 9/26-10/10, 10/10-10/17, 10/17-11/13

Condition of Sondes: 9/26-10/10 (very heavy fouling caused probe malfunctions), 10/10-10/17 (very heavy fouling), 10/17-11/13 (very heavy fouling)

Removed Data:

Parameter(s)	Problem	Data Points
All Parameters	EXP	10/15 1000-1300, 10/25 800-931, 10/26 900-1030, 10/27 930-1030, 10/29 1101-1331, 10/30 1201-1400, 10/31 1330-1500
Oxygen	FOUL	10/4 2030-10/10 1300
Oxygen	FOUL	10/14 1001-10/17
Oxygen	FOUL	10/21 331-11/13
Turbidity	FOUL	10/8 1330-10/10 1300
Turbidity	FOUL	10/25 030-11/13
Chlorophyll	FOUL	10/8 1330-10/10 1300
Chlorophyll	FOUL	10/25 030-11/13

Corrected Data:

9/26-10/10

Specific Conductivity:

standard 12.88, probe 11.8

Formula: $= -((11.8 - 12.88) / (\$B\$2113 - \$B\$1447)) * (B1447 - \$B\$1447) + E1447$

Salinity $= (0.7098 * R1447) - 2.8635$

Oxygen:

pre-deployment O2: 100.5

post-deployment O2: 6.67

**No correction possible due to fouling

Turbidity:

standard 0, probe 241.2; standard 100, probe 106.3

**No correction possible due to fouling

Chlorophyll:

standard 0.0, probe 44.5

**No correction possible due to fouling

10/10-10/17

Specific Conductivity:

standard 10, probe 11.63

**Correction not applied due to fouling (see below)

Oxygen:

pre-deployment O2: 98.87

post-deployment O2: 34.73

standard 100, probe 27.6

**No correction possible due to fouling

Turbidity:

standard 0, probe 3.5; standard 100, probe 38.4

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

**No correction possible due to fouling

Chlorophyll:

standard 0.0, probe 1.9

**No correction possible due to fouling

10/17-11/13

Specific Conductivity:

standard 12.88, probe 15.64

**Correction not applied due to fouling (see below)

Oxygen:

pre-deployment O2: 106.78

post-deployment O2: 35.22

standard 100, probe 47.7

**No correction possible due to fouling

Turbidity:

standard 0, probe 81.5; standard 100, probe 146.91

**No correction possible due to fouling

Chlorophyll:

standard 0.0, probe 4.8

**No correction possible due to fouling

Problems and Anomalies:

Oxygen 10/4 2030-10/10 1300: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Turbidity 10/8 1330-10/10 1300: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Chlorophyll 10/8 1330-10/10 1300: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Salinity 10/10: When sondes were changed on 10/10, salinity increased from 21.6ppt to 27.3ppt. This discontinuity is **faulty**.

Salinity 10/10-10/17: Sondes were fouling very badly during the entire month resulting in depression of salinity measurements over time. Between sonde deployment and recovery salinity steadily decreased, but jumped back up to higher values when a new sonde was deployed likely reflecting the slow build up of fouling organisms. If the correction was applied, the discontinuity would have been worse. As a result salinity was not corrected. The steady decrease in salinity during this period likely does not reflect actual changes in salinity and should be interpreted with caution.

Oxygen 10/14 1001-10/17: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Salinity 10/17: When sondes were changed on 10/17, salinity increased from 21ppt to 27.5ppt. This discontinuity is **faulty**.

Salinity 10/17-11/13: Sondes were fouling very badly during the entire month resulting in depression of salinity measurements over time. Between sonde deployment and recovery salinity steadily decreased, but jumped back up to higher values when a new sonde was deployed likely reflecting the slow build up of fouling organisms. If the correction was applied, the discontinuity would have been worse. As a result salinity was not corrected. The steady decrease in salinity during this period likely does not reflect actual changes in salinity and should be interpreted with caution.

Oxygen 10/21 331-11/13: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Turbidity 10/25 030-11/13: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

Chlorophyll 10/25 030-11/13: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

NOVEMBER--2003

Files: Data: DI-0311-raw, DI-0311-QAQC, DI-0311

Deployments: (2); 10/17-11/13, 11/13-12/12

Condition of Sondes: 10/17-11/13 (very heavy fouling), 11/13-12/12 (moderate fouling; multiple probe malfunctions)

Removed Data:

Parameter(s)	Problem	Data Points
All	EXP	11/1 1500-1530, 11/8 801-931, 11/9 730-1130, 11/9 2200-2231, 11/10 731-1231, 11/11 931-1130
Oxygen	FOUL	10/21 331-11/13
Oxygen	PF	11/13-12/12
Turbidity	FOUL	10/25 030-11/13
Turbidity	PF	11/13-12/12
Chlorophyll	FOUL	10/25 030-11/13
Chlorophyll	PF	11/13-12/12

Corrected Data:

10/17-11/13

Specific Conductivity:

standard 12.88, probe 15.64

**Correction not applied due to fouling (see below)

Oxygen:

pre-deployment O2: 106.78

post-deployment O2: 35.22

standard 100, probe 47.7

**No correction possible due to fouling

Turbidity:

standard 0, probe 81.5; standard 100, probe 146.91

**No correction possible due to fouling

Chlorophyll:

standard 0.0, probe 4.8

**No correction possible due to fouling

11/13-12/12

Specific Conductivity:

standard 12.88, probe 13.07

Formula: $= -((13.07 - 12.88) / (\$B\$2691 - \$B\$1291)) * (B1291 - \$B\$1291) + D1291$

Salinity $= (0.7078 * R1291) - 2.7255$

Oxygen:

**No correction possible. probe malfunction

Turbidity:

**No correction possible. probe malfunction

Chlorophyll:

**No correction possible. probe malfunction

Problems and Anomalies:

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

Salinity 10/17-11/13: Sondes were fouling very badly during the entire month resulting in depression of salinity measurements over time. Between sonde deployment and recovery salinity steadily decreased, but jumped back up to higher values when a new sonde was deployed likely reflecting the slow build up of fouling organisms. If the correction was applied, the discontinuity would have been worse. As a result salinity was not corrected. The steady decrease in salinity during this period likely does not reflect actual changes in salinity and should be interpreted with caution.

Oxygen 10/21 331-11/13: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Turbidity 10/25 030-11/13: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Chlorophyll 10/25 030-11/13: Probe was heavily fouled resulting in readings that were **faulty** and not correctable. The affected portion of the data was deleted.

Salinity 10/17: When sondes were changed on 11/13, salinity increased from 24ppt to 28ppt. This discontinuity is **faulty**.

Oxygen 11/13-12/12: Probe malfunctioned resulting in complete loss of data.

Turbidity 11/13-12/12: Probe malfunctioned resulting in complete loss of data.

Chlorophyll 11/13-12/12: Probe malfunctioned resulting in complete loss of data.

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

DECEMBER--2003

Files: Data: DI-0312-raw, DI-0312-QAQC, DI-0312

Deployments: (2); 11/13-12/12, 12/12-1/8

Condition of Sondes: 11/13-12/12 (moderate fouling; multiple probe malfunctions), 12/12-1/8 (questionable post-deployment check)

Removed Data:

Parameter(s)	Problem	Data Points
Oxygen	PF	11/13-12/12
Turbidity	PF	11/13-12/12
Chlorophyll	PF	11/13-12/12

Corrected Data:

11/13-12/12

Specific Conductivity:

standard 12.88, probe 13.07

Formula: $= -((13.07 - 12.88) / (\$B\$2691 - \$B\$1291)) * (B1291 - \$B\$1291) + D1291$

Salinity $= (0.7078 * R1291) - 2.7255$

Oxygen:

**No correction possible. probe malfunction

Turbidity:

**No correction possible. probe malfunction

Chlorophyll:

**No correction possible. probe malfunction

12/12-1/8

Specific Conductivity:

standard 12.88, probe 13.11

Formula: $= -(((13.11 - 12.88) / (\$B\$2698 - \$B\$1403)) * (B1403 - \$B\$1403)) + D1403$

Salinity $= (0.712 * R1403) - 2.95$

Oxygen:

pre-deployment O2: 95

post-deployment O2: 133.63

standard 100, probe 112.13

Formula: $= -(((132.63 - 95) / (\$B\$2698 - \$B\$1403)) * (B1403 - \$B\$1403)) + F1403 + (100 - 95)$

Conversion for O2 concentration: $= 4.91 - (0.217 * C1403) - (0.0727 * S1403) + (0.0881 * P1403)$

Turbidity:

standard 0, probe 0.4; standard 100, probe 127.2

Formula: $= (((((123 / 126.8) - 1) * (J1403)) - (0.4)) * ((B1403 - \$B\$1403) / (\$B\$2698 - \$B\$1403))) + J1403$

Chlorophyll:

standard 0.0, probe 0.5

Formula: $= -((0.5 - 0) / (\$B\$2698 - \$B\$1403)) * (B1403 - \$B\$1403) + K1403$

Problems and Anomalies:

Oxygen 11/13-12/12: Probe malfunctioned resulting in complete loss of data.

Turbidity 11/13-12/12: Probe malfunctioned resulting in complete loss of data.

Chlorophyll 11/13-12/12: Probe malfunctioned resulting in complete loss of data.

**2003 Dog Island Lease Area, Levy County
Quality Assurance/Quality Control (QA/QC) Log**

All parameters 12/12-1/8: Post-deployment check performed 12 days after sonde retrieval. Correction values may not be correct, but few are very large in magnitude so they are likely sufficient.